



**MAINTENANCE TRACKING TOOL**  
**PETTRACE800**

Date:2024-02-08

<b>Country:</b> Iberia	<b>Site:</b> BAR
<b>Intervention:</b>	Programmed maintenance: UBM/CBM <input checked="" type="checkbox"/>
<b>Subsystems:</b>	

**PRE-MAINTENANCE**

Registration Date: 2024-02-08

Gas flow(sccm): 3.5

**TPG Settings Verifications**

	Low limit (x10-)	High limit (x10-)
<b>Piranni 1 (TPG300 A1):</b>	1.00E-1	7.00E-1
<b>Piranni 2 (TPG300 A2):</b>	7.00E-2	2.00E-1
<b>Penning:</b>	1.80E-5	3.00E-5

Notes

--

Gauge number	Pressure (x10-) without gas	Pressure (x10-) with gas
<b>A1 (mbar):</b>	1.00E-3	4.60E-2
<b>A2 Under Range:</b>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<b>A2:</b>	0.00E+0	0.00E+0
<b>B1 (mbar):</b>	3.20E-8	1.50E-5

**System software**

Subsystem	Version
<b>Master:</b>	
<b>ACS:</b>	4.3.2
<b>Service System:</b>	3.6.0
<b>Manager:</b>	
<b>Informix (only applicable to SUN-Master Station):</b>	

Comments

--

Paper Burn Before PM

<b>Photos</b>
There is not photographic evidence

