

NIGHT INCIDENT



Report

Site

Barcelona

Date and time

2023-08-21 02:35

Attended

Phone Site

Department

Engineering

Scope

Yes No

System

Cyclotron - WCS

System description

Describe the incident

WCS stopped due to high temp

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

The primary filter was partially obstructed

If possible, describe the solution

If possible, describe the root cause

TotalTimeOfIntervention

20.0