

NIGHT INCIDENT



Report

Site

Barcelona

Date and time

2023-08-21 06:20

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Others - N2 gas

System description

Describe the incident

Leak in N2 line (fastlab)

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

If possible, describe the solution

Leak in fastlab module (loose connection) N2 line empty due to the leak.

If possible, describe the root cause

TotalTimeOfIntervention

5.0