

NIGHT INCIDENT



Report

Site

Oporto

Date and time

2023-09-15 01:25

Attended

Phone Site

Department

Engineering

Scope

Yes No

System

Cyclotron - IS

System description

Describe the incident

When the technician started the irradiation for the batch an error appeared "ACS Fault in ion source, gas flow fault"

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

If possible, describe the solution

Restart the irradiation

If possible, describe the root cause

TotalTimeOfIntervention

60.0