

NIGHT INCIDENT



Report

Site

Barcelona

Date and time

2023-09-09 02:58

Attended

Phone Site

Department

Engineering

Scope

Yes No

System

Others - Bunker Door

System description

Describe the incident

The operator was not able to rearm the bunker of the Petrace Door.

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

The call was longer because the operator was very anxious. I tried to calm down him.

If possible, describe the solution

The operator cleaned the sensors and reopened and closed the door again. Then he could rearm the cyclotron properly.

If possible, describe the root cause

There was a lot of dust, after the construction work in recent weeks.

TotalTimeOfIntervention

30.0