

# NIGHT INCIDENT



## Report

Site

Barcelona

Date and time

2023-09-26 05:00

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Cyclotron - WCS

System description

Describe the incident

IBA cyclotron wasnt working, it had some problems with the WCS.

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

NA

If possible, describe the solution

The WCS system had to be tested from inside the bunker.

If possible, describe the root cause

Sometimes, when the cyclotron is stop for a while we have this type of issues. We are planning to add a atomic valve purge to the system.

TotalTimeOfIntervention

60.0