

NIGHT INCIDENT



Report

Site

Murcia

Date and time

2024-01-16 07:29

Attended

Phone Site

Department

Engineering

Scope

Yes No

System

System description

Describe the incident

The QP wants to irradiate decompensated (50+80) because a dopa failed

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

If possible, describe the solution

They must irradiate compensated so they down target 4 from 80 to 50 and after to deliver stop target 1, fill target 4 again and irradiate 80+80

If possible, describe the root cause

TotalTimeOfIntervention

10.0