

NIGHT INCIDENT



Report

Site

Oporto

Date and time

2024-02-14 05:45

Attended

Phone  Site

Department

Engineering

Scope

Yes  No

System

Cyclotron - IS

System description

Describe the incident

4 times gas flow fault error

Solved?

Yes  No

Derived?

Yes  No

To whom?

Yes - Site manager  Yes - Production manager  Yes - QA  Yes - QC  Yes - QP  Others

Who is other?

Production affected?

No  Yes - Totally cancelled  Yes - Partially cancelled  Yes - Delayed

Any comment?

If possible, describe the solution

To try again

If possible, describe the root cause

We are waiting GE for the new version of MFC.

TotalTimeOfIntervention

5.0