

NIGHT INCIDENT



Report

Site

Murcia

Date and time

2024-07-02 12:38

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Cyclotron - LTF

System description

Describe the incident

The can't start the production because the error "Target pressure below limit"

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

If possible, describe the solution

Play with the valve of LTF

If possible, describe the root cause

The valve V1

TotalTimeOfIntervention

210.0