

# NIGHT INCIDENT



## Report

Site

Murcia

Date and time

2024-07-05 01:10

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Cyclotron - Extraction

System description

Describe the incident

Motor carrier 2 fault

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

If possible, describe the solution

Move the motor with service pc

If possible, describe the root cause

TotalTimeOfIntervention

20.0