

NIGHT INCIDENT

Report



Site

Oporto

Date and time

2024-07-26 01:30

Attended

Phone  Site

Department

Engineering

Scope

Yes  No

System

Others - Describe it

System description

Planning

Describe the incident

The technician by mistake was doing the irradiation at 45uA and after 30 min saw that it was planned at 65uA. The technician called to know what to do.

Solved?

Yes  No

Derived?

Yes  No

To whom?

Yes - Site manager  Yes - Production manager  Yes - QA  Yes - QC  Yes - QP  Others

Who is other?

Production affected?

No  Yes - Totally cancelled  Yes - Partially cancelled  Yes - Delayed

Any comment?

If possible, describe the solution

If possible, describe the root cause

TotalTimeOfIntervention

5.0