

# NIGHT INCIDENT

## Report



Site

Oporto

Date and time

2024-05-12 23:13

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Cyclotron - LTF

System description

Describe the incident

Air leak in LTF 2

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

If possible, describe the solution

Replacement of the V4' air valve gasket.

If possible, describe the root cause

TotalTimeOfIntervention

30.0