

NIGHT INCIDENT



Report

Site

Barcelona

Date and time

2024-07-31 00:20

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Others - Describe it

System description

Technician told me that the CFL of lab 1 was not working.

Describe the incident

We try by phone to reset the power of the electrical panel. But it doesn't work, so I went to the site. The panel was powered, but the buttons and the display weren't working properly. I checked the fuses and I tried to make it work but I couldn't.

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

When Jordi came to the site later on and could fix the problem.

If possible, describe the solution

Later Jordi find that capacitor was missed and some fuses were blown (my fault probably)

If possible, describe the root cause

The PCB was damaged

TotalTimeOfIntervention

180.0