

NIGHT INCIDENT



Report

Site

Barcelona

Date and time

2023-03-28 04:00

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Synthesizers - TRASIS AIO

System description

Describe the incident

No activity received.

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

Suspecting bad pinch-valve TBD

If possible, describe the solution

If possible, describe the root cause

TotalTimeOfIntervention

45.0