

NIGHT INCIDENT



Report

Site

Murcia

Date and time

2023-04-03 08:15

Attended

Phone Site

Department

Engineering

Scope

Yes No

System

Cyclotron - ACU

System description

Describe the incident

The computer was frozen.

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

If possible, describe the solution

Reboot the PC and the ACU

If possible, describe the root cause

TotalTimeOfIntervention

10.0