

# NIGHT INCIDENT



## Report

Site

Barcelona

Date and time

2023-05-02 05:00

Attended

Phone  Site

Department

Engineering

Scope

Yes  No

System

Cyclotron - WCS

System description

Describe the incident

WCS low pressure issues

Solved?

Yes  No

Derived?

Yes  No

To whom?

Yes - Site manager  Yes - Production manager  Yes - QA  Yes - QC  Yes - QP  Others

Who is other?

Production affected?

No  Yes - Totally cancelled  Yes - Partially cancelled  Yes - Delayed

Any comment?

If possible, describe the solution

Check for leaks

If possible, describe the root cause

TotalTimeOfIntervention

60.0