

NIGHT INCIDENT



Report

Site

Barcelona

Date and time

2023-05-30 05:00

Attended

Phone

Site

Department

Engineering

Scope

Yes

No

System

Distribution - Peristaltic pump

System description

Describe the incident

No communication with the peristaltic pump

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

USB to RS232 comm issues

If possible, describe the solution

TBD

If possible, describe the root cause

TotalTimeOfIntervention

30.0