

NIGHT INCIDENT



Report

Site

Oporto

Date and time

2023-06-01 00:57

Attended

Phone Site

Department

Engineering

Scope

Yes No

System

System description

Describe the incident

The technician started the irradiation with the wrong target

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

If possible, describe the solution

Replanned with the target 1 and don't used the target 4 because the batch was very loaded and the target 4 is very worn

If possible, describe the root cause

TotalTimeOfIntervention

30.0