

NIGHT INCIDENT

Report



Site

Meldola

Date and time

2023-05-06 09:14

Attended

Phone Site

Department

QP

Scope

Yes No

System

System description

Describe the incident

Leak test non passata per mancanza accensione compressore

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

If possible, describe the solution

If possible, describe the root cause

TotalTimeOfIntervention

15.0