

NIGHT INCIDENT



Report

Site

Venafro

Date and time

2023-05-22 23:06

Attended

Phone

Site

Department

E&S

Scope

Yes

No

System

System description

Describe the incident

Gas electronic valve triggered by alarm

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

If possible, describe the solution

Bypass on the system

If possible, describe the root cause

TotalTimeOfIntervention

10.0