

# NIGHT INCIDENT



## Report

Site

Meldola

Date and time

2023-10-03 00:15

Attended

Phone

Site

Department

QC

Scope

Yes

No

System

System description

Describe the incident

pHmeter Memory doesn't work, technician needs instruction on how to manager It.

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

If possible, describe the solution

pHmeter PH05 replaced with PH06

If possible, describe the root cause

TotalTimeOfIntervention

15.0