

# NIGHT INCIDENT



## Report

Site

Venafro

Date and time

2023-06-04 16:45

Attended

Phone

Site

Department

Production,E&S

Scope

Yes

No

System

System description

Describe the incident

HVV closed due to bad value in B1 sensor

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

We had trouble finding the spare part we needed

If possible, describe the solution

To be sure to solve the issue we have to replace the penning cable (SHV-MHV, 40 m, 1500V, 75 ohm). At the moment we don't have spare parts so we can only cut 30 cm in the end of the cable and fix the connector cable-penning as well as possible

If possible, describe the root cause

Penning cable damaged due to radiation inside the bunker

TotalTimeOfIntervention

2880.0