

NIGHT INCIDENT

Report



Site

Meldola

Date and time

2024-08-02 00:17

Attended

Phone Site

Department

Production

Scope

Yes No

System

System description

Describe the incident

Allarme flusso CP06

Solved?

Yes No

Derived?

Yes No

To whom?

Yes - Site manager Yes - Production manager Yes - QA Yes - QC Yes - QP Others

Who is other?

Production affected?

No Yes - Totally cancelled Yes - Partially cancelled Yes - Delayed

Any comment?

If possible, describe the solution

Na

If possible, describe the root cause

Spegnimento IPM prima dello spegnimento del contaparticelle

TotalTimeOfIntervention

15.0