

# NIGHT INCIDENT



## Report

Site

Venafro

Date and time

2024-08-06 05:10

Attended

Phone  Site

Department

E&S,Production

Scope

Yes  No

System

System description

Describe the incident

Issue in activity transfer from synthesizer to dispensing system

Solved?

Yes  No

Derived?

Yes  No

To whom?

Yes - Site manager  Yes - Production manager  Yes - QA  Yes - QC  Yes - QP  Others

Who is other?

Production affected?

No  Yes - Totally cancelled  Yes - Partially cancelled  Yes - Delayed

Any comment?

93 GBq instead of 434 GBq arrived in the dispensing system

If possible, describe the solution

Activity transfer forced manually in maintenance mode

If possible, describe the root cause

To investigate asap

TotalTimeOfIntervention

30.0