

NIGHT INCIDENT



Report

Site

Meldola

Date and time

2023-08-01 04:35

Attended

Phone

Site

Department

Production

Scope

Yes

No

System

System description

Describe the incident

Errori multipli su bilancia ID03. No communication, Er 243-290

Solved?

Yes

No

Derived?

Yes

No

To whom?

Yes - Site manager

Yes - Production manager

Yes - QA

Yes - QC

Yes - QP

Others

Who is other?

Production affected?

No

Yes - Totally cancelled

Yes - Partially cancelled

Yes - Delayed

Any comment?

Ritardo per cambio schermo eseguito dopo misura madre, non ci sono altri impatti sul lotto.

If possible, describe the solution

Sostituito schermo esterno

If possible, describe the root cause

TotalTimeOfIntervention

30.0